


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2065, Rack A, Plate 1, Well B11, 490392
[CD3734BF9A5B7317] Feb 04, 2026 12:22:42 Pressure = 86 bar (85 to 88)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.507	F 0.61	0.8%	9738	
2	0.697	F 0.83	15.0%	182885	Acetylated F peak
3	0.832	F 1.00	75.7%	924415	Consistent with F
4	1.190	A 1.01	8.6%	104594	A peak
Total Area: 1221632					

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range