


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2188, Rack A, Plate 1, Well D02, 502182
[C939C4C32ADD91B5] Jul 01, 2026 12:09:19 Pressure = 79 bar (79 to 81)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.520	F 0.62	0.7%	9792	
2	0.693	F 0.83	10.9%	156963	Acetylated F peak
3	0.825	F 0.99	67.8%	980034	Consistent with F
4	1.167	A 1.01	20.6%	298315	A peak - REVIEW
Total Area:				1445104	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range