


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2205, Rack A, Plate 1, Well C03, 496031
[DC2DCDFD6715391F] Jul 23, 2026 13:54:08 Pressure = 95 bar (94 to 97)

AF



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.423	F	0.23	1.5%	23388	
2	0.523	F	0.28	2.0%	32627	
3	0.805	F	0.43	1.1%	17975	
4	1.235	F	0.66	1.4%	21927	
5	1.877	F	1.00	2.3%	37177	6
6	2.592	F	1.39	7.9%	126194	3
7	3.440	A	0.99	81.5%	1306995	A peak
8	4.352	S	0.86	2.4%	38253	

Total Area: 1604536

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range