


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2043, Rack A, Plate 1, Well A03, 171834-P
[C873C8332EC206EA] Jan 15, 2026 15:12:28 Pressure = 82 bar (81 to 84)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.402	F 0.48	1.5%	12735	
2	0.502	F 0.59	0.5%	4606	
3	0.697	F 0.82	15.1%	130474	Acetylated F peak
4	0.838	F 0.99	66.7%	575092	Consistent with F
5	1.205	A 1.01	16.2%	139642	A peak
Total Area:				862549	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range