


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 1865, Rack A, Plate 1, Well A03, 770583
 [CC60DE8F6A59B5] Jul 16, 2025 10:24:15 Pressure = 61 bar (59 to 61)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.397	F 0.47	1.9%	14727	
2	0.493	F 0.59	1.3%	10447	
3	0.695	F 0.83	15.8%	123747	Acetylated F peak
4	0.837	F 1.00	59.2%	463219	Consistent with F
5	0.995	F 1.18	2.3%	18206	
6	1.232	A 1.02	19.0%	149020	A peak
7	1.375	A 1.14	0.4%	3070	
Total Area:				782436	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range