


LABORATORIO PREGEN  
 Carrera 15a No 106-42  
 BOGOTA

Batch 2022, Rack A, Plate 1, Well A12, 484370  
 [D973D1D322F55B9D] Dec 19, 2025 10:28:09 Pressure = 76 bar (75 to 77)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.402	F 0.47	1.0%	8761	
2	0.497	F 0.59	0.8%	6878	
3	0.708	F 0.83	13.1%	113225	Acetylated F peak
4	0.848	F 1.00	66.4%	575924	Consistent with F
5	0.993	F 1.17	1.3%	10869	
6	1.227	A 1.02	17.5%	151483	A peak
Total Area:				867140	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range