


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 1948, Rack A, Plate 1, Well C10, 889803  
[89354D5AEB3CF13F] Oct 15, 2025 11:04:04

Pressure = 68 bar (67 to 69)

FA



| PEAK                | RT    | REL | RT   | % CONC | AREA    | COMMENT           |
|---------------------|-------|-----|------|--------|---------|-------------------|
| 1                   | 0.493 | F   | 0.60 | 5.7%   | 81798   |                   |
| 2                   | 0.818 | F   | 0.99 | 77.3%  | 1110778 | Consistent with F |
| 3                   | 1.178 | A   | 1.00 | 17.0%  | 244688  | A peak            |
| Total Area: 1437264 |       |     |      |        |         |                   |

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) A1c > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range