


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 1974, Rack A, Plate 1, Well C06, 170980
 [9C62A3EE67F8A44A] Nov 11, 2025 12:10:08 Pressure = 66 bar (66 to 68)

AF



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.388	F 0.45	2.4%	15270	
2	0.487	F 0.56	2.3%	14497	
3	0.718	F 0.83	6.2%	39980	Acetylated F peak
4	0.872	F 1.01	31.4%	201757	Consistent with F
5	1.012	F 1.17	2.5%	16361	
6	1.102	A 0.89	0.5%	3431	
7	1.245	A 1.00	54.6%	350825	A peak - REVIEW
Total Area:				642121	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range