


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2127, Rack A, Plate 1, Well E10, 173027-P  
[8D23098AC2CBF93F] Apr 21, 2026 11:41:56 Pressure = 62 bar (62 to 64)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.503	F 0.58	9.3%	51695	
2	0.720	F 0.83	18.0%	100327	
3	0.865	F 1.00	57.3%	318688	Consistent with F
4	1.238	A 1.00	15.3%	85333	A peak

Total Area: 556043

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range