


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1936, Rack A, Plate 1, Well F08, 170540
[8C7049C897945B1F] Oct 08, 2025 13:18:40 Pressure = 65 bar (64 to 66)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.400	F 0.47	1.6%	9959	
2	0.495	F 0.58	1.0%	6215	
3	0.692	F 0.82	13.3%	84810	Acetylated F peak
4	0.828	F 0.98	57.6%	366466	Consistent with F
5	0.968	F 1.14	2.0%	12840	
6	1.183	A 1.00	24.5%	155609	A peak - REVIEW
Total Area:				635899	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range