


LABORATORIO PREGEN  
 Carrera 15a No 106-42  
 BOGOTA

Batch 2137, Rack A, Plate 1, Well B03, 503256  
 [D8660BBF9A968468] Apr 30, 2026 12:00:37 Pressure = 63 bar (62 to 64)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.427	F 0.50	1.7%	24410	
2	0.513	F 0.60	0.8%	11117	
3	0.717	F 0.84	19.1%	277898	Acetylated F peak
4	0.870	F 1.02	66.1%	963252	Consistent with F
5	1.220	A 1.00	12.4%	180190	A peak
Total Area:				1456867	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range