


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2182, Rack A, Plate 1, Well D03, 486257
[997DEA2DC43A4EEA] Jun 23, 2026 11:27:56 Pressure = 73 bar (72 to 74)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.518	F 0.61	1.0%	16093	
2	0.698	F 0.82	15.9%	254337	Acetylated F peak
3	0.825	F 0.97	75.2%	1207281	Consistent with F
4	1.180	A 1.01	7.9%	126750	A peak
Total Area:				1604461	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range