


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2188, Rack A, Plate 1, Well C11, 501476  
[CD3D91967F8A6C6A] Jul 01, 2026 12:03:16 Pressure = 79 bar (79 to 82)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.517	F 0.62	0.9%	13111	
2	0.717	F 0.86	12.6%	178815	Acetylated F peak
3	0.818	F 0.98	74.8%	1064598	Consistent with F
4	1.157	A 1.00	11.8%	167548	A peak
Total Area:				1424072	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range