


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 1988, Rack A, Plate 1, Well C04, 482022
 [C833A8E08D3A0660] Nov 21, 2025 11:32:39 Pressure = 70 bar (69 to 71)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.408	F 0.48	0.6%	9208	
2	0.493	F 0.58	0.7%	10250	
3	0.713	F 0.83	10.2%	147302	Acetylated F peak
4	0.848	F 0.99	59.4%	857484	Consistent with F
5	1.002	F 1.17	1.5%	20977	
6	1.233	A 1.01	27.6%	399364	A peak - REVIEW
Total Area:				1444585	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range