


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2191, Rack A, Plate 1, Well D09, 493066  
[8938847FB3A0EEE2] Jul 06, 2026 13:11:26 Pressure = 79 bar (79 to 81)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.683	F 0.82	22.0%	524352	
2	0.813	F 0.98	63.3%	1509554	Consistent with F
3	1.173	A 1.00	14.7%	351201	A peak

Total Area: 2385107

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range