


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2022, Rack A, Plate 1, Well F04, 489901
[8837C1C2DF8A8CE2] Dec 19, 2025 12:17:18 Pressure = 74 bar (74 to 76)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.402	F 0.47	0.7%	10347	
2	0.502	F 0.58	0.6%	9401	
3	0.717	F 0.83	15.4%	230264	Acetylated F peak
4	0.847	F 0.99	69.3%	1036614	Consistent with F
5	1.225	A 1.01	14.0%	209918	A peak
Total Area:				1496544	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range