


- - - - -
`LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2182, Rack A, Plate 1, Well C02, 493536
[9D2DBE7893ED3397] Jun 23, 2026 10:59:38

FA
Pressure = 73 bar (73 to 75)



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.430	F	0.52	0.5%	6278	
2	0.520	F	0.62	0.3%	3590	
3	0.700	F	0.84	17.6%	214347	Acetylated F peak
4	0.822	F	0.99	76.2%	929497	Consistent with F
5	1.188	A	1.01	5.4%	66287	A peak
Total Area: 1219999						

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range