


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1919, Rack A, Plate 1, Well F03, 478610
[9870039400098E42] Sep 24, 2025 12:59:30 Pressure = 60 bar (60 to 61)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.407	F 0.48	1.2%	13770	
2	0.490	F 0.58	1.0%	10859	
3	0.703	F 0.83	12.8%	144696	Acetylated F peak
4	0.848	F 1.00	61.7%	696696	Consistent with F
5	1.200	A 1.00	23.3%	262725	A peak - REVIEW
Total Area:				1128746	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range