


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2065, Rack A, Plate 1, Well D04, 498914  
[997665EACF8EA6C8] Feb 04, 2026 12:59:05 Pressure = 87 bar (86 to 89)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.507	F 0.60	1.6%	19196	
2	0.707	F 0.84	21.5%	256507	Acetylated F peak
3	0.828	F 0.98	73.0%	870545	Consistent with F
4	1.203	A 1.02	3.9%	46004	A peak

Total Area: 1192252

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range