


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 1965, Rack A, Plate 1, Well A04, 482137
 [CC37B0870A7AA4CA] Oct 31, 2025 13:50:32 Pressure = 68 bar (65 to 68)

AF



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.385	F 0.19	3.4%	38792	
2	0.522	F 0.26	3.1%	35497	
3	0.900	F 0.45	9.5%	109366	
4	1.698	F 0.85	0.6%	6847	
5	1.978	F 0.99	21.9%	252253	Consistent with F
6	2.750	F 1.38	3.3%	38434	
7	2.970	A 0.79	2.1%	23733	
8	3.243	A 0.86	0.4%	4570	
9	3.752	A 1.00	54.3%	625235	A peak - REVIEW
10	4.267	A 1.13	0.2%	2733	
11	4.510	A 1.20	1.2%	13826	
Total Area:				1151286	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range