


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 1791, Rack A, Plate 1, Well C03, 457122
 [C924E32ACE070CC2] May 07, 2025 10:54:02 Pressure = 59 bar (58 to 60)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.403	F 0.49	1.1%	10447	
2	0.500	F 0.60	0.5%	4130	
3	0.718	F 0.87	16.9%	153489	Acetylated F peak
4	0.833	F 1.00	74.5%	677688	Consistent with F
5	1.210	A 1.01	7.0%	63549	A peak
Total Area:				909303	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range