


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2173, Rack A, Plate 1, Well A12, 502150
[C879EC72D390C442] Jun 10, 2026 13:23:46 Pressure = 68 bar (67 to 69)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.517	F 0.62	1.1%	22827	
2	0.695	F 0.83	15.5%	317227	Acetylated F peak
3	0.823	F 0.98	72.2%	1476494	Consistent with F
4	1.192	A 1.00	11.2%	229584	A peak
Total Area:				2046132	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range