


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1948, Rack A, Plate 1, Well A09, 170597
[D8251C4AC19CD315] Oct 15, 2025 10:11:31 Pressure = 67 bar (67 to 69)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.497	F 0.60	1.7%	16107	
2	0.688	F 0.83	15.2%	144562	Acetylated F peak
3	0.815	F 0.98	67.7%	642412	Consistent with F
4	0.975	F 1.18	1.4%	12897	
5	1.190	A 1.01	14.1%	133488	A peak
Total Area:				949466	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range