


LABORATORIO PREGEN  
 Carrera 15a No 106-42  
 BOGOTA

Batch 2127, Rack A, Plate 1, Well G01, 502959  
 [89225DDF951E26CA] Apr 21, 2026 12:14:16 Pressure = 62 bar (62 to 63)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.430	F 0.50	1.7%	19533	
2	0.505	F 0.58	0.9%	9848	
3	0.740	F 0.85	15.6%	178508	Acetylated F peak
4	0.855	F 0.99	64.2%	732346	Consistent with F
5	1.247	A 1.01	17.6%	201130	A peak
Total Area:				1141365	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range