


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2169, Rack A, Plate 1, Well A10, 501546
[993DFDEB186D933F] Jun 05, 2026 13:29:56 Pressure = 65 bar (65 to 67)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.512	F	0.61	1.8%	22739	
2	0.703	F	0.84	14.7%	184188	Acetylated F peak
3	0.838	F	1.00	74.1%	927293	Consistent with F
4	1.185	A	1.00	9.3%	116560	A peak
Total Area: 1250780						

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range