


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2098, Rack A, Plate 1, Well B10, 498825
[DC327A5403C6846A] Mar 19, 2026 11:56:20 Pressure = 60 bar (60 to 62)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.425	F 0.49	1.8%	16972	
2	0.512	F 0.59	0.6%	5813	
3	0.713	F 0.82	17.7%	169497	Acetylated F peak
4	0.857	F 0.99	61.4%	586508	Consistent with F
5	1.007	F 1.16	2.1%	19633	
6	1.248	A 1.02	16.5%	157292	A peak

Total Area: 955715

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range