


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2182, Rack A, Plate 1, Well G03, 486202
[993DBF3DEE38CC40] Jun 23, 2026 12:42:43 Pressure = 71 bar (71 to 73)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.515	F 0.61	1.7%	47202	
2	0.702	F 0.83	17.0%	485313	Acetylated F peak
3	0.832	F 0.99	68.3%	1954038	Consistent with F
4	1.167	A 1.00	13.1%	373630	A peak
Total Area:				2860183	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range