


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2194, Rack A, Plate 1, Well D04, 493073
[886C844AEBA2E66A] Jul 09, 2026 11:40:04

Pressure = 80 bar (80 to 83)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.682	F	0.82	20.1%	268133	
2	0.817	F	0.99	66.3%	883201	Consistent with F
3	1.157	A	1.00	13.5%	180156	A peak
Total Area: 1331490						

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range