


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1824, Rack A, Plate 1, Well A10, 456858
[9D65C0E427A87937] Jun 04, 2025 10:59:18 Pressure = 57 bar (57 to 59)

FA



| PEAK | RT | REL RT | % CONC | AREA | COMMENT |
|-------------|-------|--------|--------|---------|-------------------|
| 1 | 0.398 | F 0.48 | 1.1% | 12183 | |
| 2 | 0.493 | F 0.59 | 1.1% | 11329 | |
| 3 | 0.725 | F 0.87 | 12.9% | 138699 | Acetylated F peak |
| 4 | 0.845 | F 1.01 | 61.2% | 659410 | Consistent with F |
| 5 | 1.218 | A 1.01 | 23.7% | 255289 | A peak - REVIEW |
| Total Area: | | | | 1076910 | |

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range