


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2204, Rack A, Plate 1, Well D07, 496047  
[C8289DBC3A4066C8] Jul 23, 2026 11:38:01 Pressure = 93 bar (93 to 95)

FA



| PEAK                | RT    | REL | RT   | % CONC | AREA   | COMMENT           |
|---------------------|-------|-----|------|--------|--------|-------------------|
| 1                   | 0.698 | F   | 0.84 | 19.4%  | 222327 | Acetylated F peak |
| 2                   | 0.830 | F   | 1.00 | 70.3%  | 805787 | Consistent with F |
| 3                   | 1.168 | A   | 1.01 | 10.3%  | 118050 | A peak            |
| Total Area: 1146164 |       |     |      |        |        |                   |

Codes: 1) Wide A peak  
2) Area of A peak < 80%  
3) Peak area greater than expected  
4) Peak after A2  
5) A1c > 10%  
6) HbF or variant present  
7) Total sample area too small/big  
8) A2 is not within normal range