


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1818, Rack A, Plate 1, Well B02, 455892
[CD64FECB1747A44A] May 28, 2025 09:51:28 Pressure = 61 bar (61 to 62)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.403	F 0.48	2.5%	11953	
2	0.485	F 0.57	4.4%	21373	
3	0.715	F 0.84	5.4%	26314	Acetylated F peak
4	0.857	F 1.01	46.3%	225549	Consistent with F
5	0.972	F 1.15	3.2%	15475	
6	1.233	A 1.02	38.1%	185789	A peak - REVIEW
7	1.392	S 0.89	0.2%	1038	
Total Area:				487491	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range