


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2173, Rack A, Plate 1, Well E12, 503033
 [9C6CB862FB106648] Jun 10, 2026 15:04:49

Pressure = 67 bar (66 to 68) FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.428	F 0.50	0.9%	16511	
2	0.510	F 0.60	0.3%	6545	
3	0.697	F 0.82	14.9%	281868	Acetylated F peak
4	0.827	F 0.97	68.7%	1300864	Consistent with F
5	1.195	A 1.00	15.2%	287787	A peak
Total Area:				1893575	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range