


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1361, Rack A, Plate 1, Well A02, 27406
[CD9DA4438127373F] Apr 03, 2024 15:04:52

Pressure = 61 bar (61 to 61)

FA



| PEAK | RT | REL RT | % CONC | AREA | COMMENT |
|-------------|-------|--------|--------|--------|-------------------|
| 1 | 0.415 | F 0.47 | 2.3% | 18290 | |
| 2 | 0.485 | F 0.54 | 4.0% | 31738 | |
| 3 | 0.753 | F 0.85 | 10.8% | 85336 | Acetylated F peak |
| 4 | 0.882 | F 0.99 | 60.5% | 478861 | Consistent with F |
| 5 | 1.018 | F 1.14 | 1.9% | 15249 | |
| 6 | 1.255 | A 1.00 | 20.5% | 162546 | A peak - REVIEW |
| Total Area: | | | | 792020 | |

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range