


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2080, Rack A, Plate 1, Well D08, 490063  
[DC336312FC242CE2] Feb 24, 2026 11:52:46 Pressure = 56 bar (56 to 58)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.402	F 0.46	2.6%	22130	
2	0.505	F 0.57	1.8%	15252	
3	0.730	F 0.83	11.3%	97865	Acetylated F peak
4	0.883	F 1.00	55.6%	481181	Consistent with F
5	1.265	A 1.01	28.8%	249521	A peak - REVIEW
Total Area:				865949	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range