


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2185, Rack A, Plate 1, Well D04, 486219
[8D69AB0323E739B7] Jun 25, 2026 13:07:26 Pressure = 73 bar (72 to 74)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.510	F 0.61	1.2%	29046	
2	0.685	F 0.82	15.2%	374278	Acetylated F peak
3	0.817	F 0.98	67.0%	1653550	Consistent with F
4	1.188	A 1.01	16.6%	410346	A peak
Total Area:				2467220	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range