


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 1998, Rack A, Plate 1, Well A06, 171204  
[D926BD67AE67D397] Nov 28, 2025 10:58:34 Pressure = 72 bar (72 to 73)

FA



| PEAK        | RT    | REL RT | % CONC | AREA   | COMMENT           |
|-------------|-------|--------|--------|--------|-------------------|
| 1           | 0.388 | F 0.45 | 3.4%   | 17516  |                   |
| 2           | 0.485 | F 0.57 | 2.6%   | 13350  |                   |
| 3           | 0.708 | F 0.83 | 7.1%   | 36888  | Acetylated F peak |
| 4           | 0.865 | F 1.01 | 41.6%  | 216507 | Consistent with F |
| 5           | 1.008 | F 1.17 | 2.5%   | 13025  |                   |
| 6           | 1.245 | A 1.01 | 42.8%  | 222766 | A peak - REVIEW   |
| Total Area: |       |        |        | 520052 |                   |

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range