


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2185, Rack A, Plate 1, Well B06, 486216
[8969FA56761064C8] Jun 25, 2026 12:20:57 Pressure = 72 bar (72 to 74)

FA



| PEAK | RT | REL RT | % CONC | AREA | COMMENT |
|-------------|-------|--------|--------|---------|-------------------|
| 1 | 0.512 | F 0.61 | 1.0% | 11980 | |
| 2 | 0.688 | F 0.82 | 13.9% | 160793 | Acetylated F peak |
| 3 | 0.837 | F 1.00 | 69.0% | 796391 | Consistent with F |
| 4 | 1.187 | A 1.01 | 16.0% | 185242 | A peak |
| Total Area: | | | | 1154406 | |

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range