


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 1964, Rack A, Plate 1, Well G10, 492294
 [8D27A190A0F00CEA] Oct 31, 2025 13:25:32 Pressure = 66 bar (65 to 67)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.393	F 0.45	1.3%	13517	
2	0.495	F 0.57	0.6%	6380	
3	0.718	F 0.82	8.2%	87055	Acetylated F peak
4	0.853	F 0.98	76.5%	811554	Consistent with F
5	1.240	A 1.01	13.4%	142431	A peak
				Total Area: 1060937	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range