


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2068, Rack A, Plate 1, Well F10, 489333  
[9D33202E66F159BF] Feb 10, 2026 12:47:08 Pressure = 94 bar (94 to 97)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.505	F 0.60	1.4%	16594	
2	0.703	F 0.84	17.3%	199338	Acetylated F peak
3	0.855	F 1.02	71.4%	825305	Consistent with F
4	1.193	A 1.00	9.9%	114276	A peak
Total Area:				1155513	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range