


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2197, Rack A, Plate 1, Well E10, 487119
[DD6CC3CAE880E4E2] Jul 15, 2026 14:14:06 Pressure = 84 bar (84 to 86)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.515	F	0.62	1.3%	14602	
2	0.683	F	0.82	14.8%	167820	Acetylated F peak
3	0.818	F	0.99	72.0%	818728	Consistent with F
4	1.153	A	0.99	12.0%	136020	A peak
Total Area: 1137170						

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range