


LABORATORIO PREGEN  
 Carrera 15a No 106-42  
 BOGOTA

Batch 1905, Rack A, Plate 1, Well A05, 169998  
 [CC7041E2D25CD33D] Sep 10, 2025 11:04:53 Pressure = 62 bar (62 to 63)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.407	F 0.49	1.0%	6844	
2	0.497	F 0.60	1.3%	8609	
3	0.690	F 0.83	11.8%	79989	Acetylated F peak
4	0.830	F 1.00	61.3%	415218	Consistent with F
5	0.957	F 1.15	1.7%	11262	
6	1.195	A 1.01	22.9%	155334	A peak - REVIEW
Total Area:				677256	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range