


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2083, Rack A, Plate 1, Well A09, 506655
 [993378F78599DBB5] Feb 27, 2026 10:34:39 Pressure = 56 bar (55 to 57)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.410	F 0.47	2.0%	9753	
2	0.508	F 0.58	2.6%	12674	
3	0.730	F 0.83	14.6%	70848	Acetylated F peak
4	0.892	F 1.01	58.7%	285573	Consistent with F
5	1.113	A 0.89	0.3%	1272	
6	1.270	A 1.01	21.9%	106775	A peak - REVIEW
Total Area:				486895	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range