


LABORATORIO PREGEN  
 Carrera 15a No 106-42  
 BOGOTA

Batch 1818, Rack A, Plate 1, Well D11, 457150  
 [9C65FE9E40105B95] May 28, 2025 11:00:11 Pressure = 61 bar (60 to 62)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.410	F 0.47	0.9%	15366	
2	0.493	F 0.57	1.1%	20514	
3	0.737	F 0.85	16.3%	291766	Acetylated F peak
4	0.857	F 0.99	71.2%	1276485	Consistent with F
5	1.222	A 1.01	10.5%	187940	A peak
Total Area:				1792071	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range