


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2194, Rack A, Plate 1, Well D06, 493128
[8D79844AEBA0EC62] Jul 09, 2026 11:44:06 Pressure = 81 bar (81 to 83)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.528	F	0.64	0.8%	12376	
2	0.705	F	0.85	14.2%	221483	Acetylated F peak
3	0.813	F	0.98	74.6%	1159891	Consistent with F
4	1.152	A	1.00	10.4%	161411	A peak
Total Area: 1555161						

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range