


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2086, Rack A, Plate 1, Well F07, 172563
 [C82729717139D9BD] Mar 04, 2026 13:10:37 Pressure = 52 bar (52 to 54)

AF



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.402	F 0.45	3.6%	18148	
2	0.493	F 0.55	2.2%	10955	
3	0.733	F 0.82	2.6%	13154	
4	0.902	F 1.01	17.9%	89471	Consistent with F
5	1.030	F 1.15	3.1%	15332	
6	1.133	A 0.89	0.6%	2984	
7	1.260	A 0.99	70.0%	350135	A peak - REVIEW
Total Area:				500179	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range