


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2041, Rack A, Plate 1, Well E03, 488984
 [C937C9765315F3B7] Jan 15, 2026 12:29:18 Pressure = 82 bar (81 to 84)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.402	F 0.47	0.9%	10232	
2	0.500	F 0.58	0.5%	5953	
3	0.710	F 0.83	13.4%	157072	Acetylated F peak
4	0.830	F 0.97	54.6%	640718	Consistent with F
5	0.995	F 1.16	1.6%	18910	
6	1.205	A 1.01	28.6%	335164	A peak - REVIEW
7	1.378	S 0.88	0.4%	4755	
Total Area:				1172804	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range