


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1967, Rack A, Plate 1, Well E01, 170919
[C977A12ACCF80648] Nov 05, 2025 13:53:21 Pressure = 64 bar (64 to 66)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.388	F 0.45	5.8%	18727	
2	0.490	F 0.56	4.7%	15271	
3	0.733	F 0.84	19.2%	61939	
4	0.890	F 1.02	43.8%	140975	Consistent with F
5	1.013	F 1.16	4.1%	13184	
6	1.237	A 0.99	22.3%	71730	A peak - REVIEW

Total Area: 321826

* * * The total peak area is too low/high for evaluation! * * *

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range