


LABORATORIO PREGEN  
 Carrera 15a No 106-42  
 BOGOTA

Batch 1948, Rack A, Plate 1, Well C09, 859802  
 [DD614D5AEB3C71BF] Oct 15, 2025 11:02:04 Pressure = 68 bar (68 to 69)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.497	F 0.60	2.8%	26701	
2	0.688	F 0.83	19.0%	183693	Acetylated F peak
3	0.817	F 0.99	66.7%	644071	Consistent with F
4	0.975	F 1.18	1.4%	13596	
5	1.187	A 1.00	10.0%	96960	A peak
Total Area:				965021	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range