


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1930, Rack A, Plate 1, Well E11, 478804
[8C3458A15A6B2CC0] Oct 03, 2025 12:14:03 Pressure = 62 bar (62 to 63)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.490	F 0.58	3.7%	46284	
2	0.703	F 0.83	14.2%	178538	Acetylated F peak
3	0.853	F 1.01	60.5%	759823	Consistent with F
4	1.195	A 1.01	21.6%	272047	A peak - REVIEW
Total Area:				1256692	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range