


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 1884, Rack A, Plate 1, Well C04, 169640  
[CD717715D4A47335] Aug 13, 2025 11:37:59 Pressure = 61 bar (60 to 61)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.400	F 0.48	1.1%	8749	
2	0.495	F 0.59	0.9%	7435	
3	0.702	F 0.84	13.8%	112407	Acetylated F peak
4	0.848	F 1.01	64.2%	523881	Consistent with F
5	1.213	A 1.01	20.0%	163448	A peak
Total Area:				815920	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range